ulma International GmbH
Alfred-Nobel-Str. 5
50226 Frechen
Deutschland / Germany

TECHNICAL DATA SHEET



Reference: **Maxitex Duplex PF**

Article No.: **1121001, 1121002, 1121003, 1121004, 1121005, 1121006, 1121007,**

Description: Latex Surgical glove, sterile, powder-free, brown / green, category III, class IIa

Usage: For single use in the medical sector.

CHARACTERISTICS

Thickness (single wall)

	Length*	Cuff	Palm	Finger
All sizes:	min. 290mm	0,18±0,02mm	0,21±0,02mm	0,24±0,02mm

	6,0	6,5	7,0	7,5	8,0	8,5	9,0
Palm width:	72±0,05mm	77±0,05mm	83±0,05mm	89±0,05mm	95±0,05mm	102±0,05mm	108±0,05mm

**in accordance to EN 455-2*

DESIGN

Material:	latex	Color:	brown / green
Powder:	powder-free	External surface:	textured
Powder content**:	max. 2 mg per glove	Shape:	handspecific
Protein content***:	≤ 50 µg/g	Cuff:	beaded cuff
Internal surface:	polymer-coated	Sterilization:	sterile (Gamma rays or electron beams)

***in accordance to ASTM D-6124 // *** in accordance to ASTM D-5712*

PHYSICAL PROPERTIES

	Before Aging		After Aging
Force at break:	min. 9 N	Force at break:	min. 9 N
AQL:	0,65		

LOGISTIC DATA

	Pouch	Dispenser	Outer carton
Quantities:	1 set	25 sets of gloves	4 x 25 sets of surgical gloves
Dimensions (LxWxH):	25,3 x 13 cm	27 x 14 x 25 cm	54,8 x 14,7 x 50,8 cm

Shelf life: 5 years

Storage: Store in original packaging in a dry and dark place at room temperature.
Protect from heat, humidity, bright light and ozone.

LABELLING

Batch number: Depending on the version
Manufacturing date: Depending on the version
Expiry date: Depending on the version

QUALITY MANAGEMENT

Product-related standards: | EN ISO 374-1:2016+A1:2018 / Type B | EN 374-2:2014 | EN 374-4:2019 | EN 374-5:2016 |
| EN 16523-1:2015+A1:2018 | EN ISO 21420:2020 | EN 421:2010 |
| EN 455-1:2020+A2:2024 | EN 455-2:2024 | EN 455-3:2023 | EN 455-4:2009 |
| EN ISO 11137-1:2015/A2:2019 |

Quality certificates: ISO 13485:2016

All errors and printing errors reserved.

ulma International GmbH
Alfred-Nobel-Str. 5
50226 Frechen
Deutschland / Germany

Medical Device:	Medical device, class IIa, in accordance to Regulation 2017/745
Personal Protective Equipment:	Personal protective equipment in accordance to Regulation (EU) 2016/425 CE Category III: For irreversible or mortal risks

CE 2777



DISPOSAL

Dispose in accordance with the valid regulations. Gloves contaminated with chemical substances must be disposed of in accordance with the regulations for the relevant chemicals.

NOTICES

Usage: The product needs to be checked for the correct size and suitability with regard to its intended use.
Before use check for damage and do not use damaged or unsuitable gloves.
The gloves are intended for single use only and shall be disposed on a regular basis.

Caution: The product may cause allergic reactions on sensitive people.
Particular caution must be exercised if the user has any known or suspected hypersensitivity.

The information above are to be considered as typicial values in the standard version based on the manufacturer's information. Subject to modification.

The gloves shall only be used for its intended purpose. If needed we're glad to support you with further details.

Technical data sheet issued by

ulma International GmbH
Alfred-Nobel-Str. 5
50226 Frechen
Deutschland / Germany

ulma
Alfred-Nobel-Str. 5
50226 Frechen
info@ulma.de

Issuing date: 10.12.2025
Expiry date: 05.08.2030

(Signature and company stamp)

This document is the property of ulma International GmbH. Any form of distributing and unauthorized copying are strictly forbidden.

All errors and printing errors reserved.